

Android application wound assessment System for Patients with Diabetes

R. Tamilselvi¹, S. Subathra²

Department of Computer Science and Engineering, Velammal Institute of Technology, Chennai, India.

Abstract: Nowadays many doctors and nurses assess the wound just by examining the size of wound and clinical status, while the patient themselves can assess their wounds. It allows a patient to take care of themselves and perform their active role in taking care on their own. It allows a patient to take more quantitative effect and cost examination and wound care potential to handle the wound effectively. It helps in saving cost and travelling expenses. As the evolution of smartphone is highly developing, it possesses a high resolution camera that helps in assessing and analysing wounds, is an attractive option. In this paper, we propose an idea of implementing image analysis system on android phone. The wound is captured on smartphone camera with the help of image capture box. Which will perform segmentation of wound with accelerated mean shift algorithm. Based on the skin color the outline of the foot is determined. The boundary of wound is found using simple connected region detection method. Next the status of healing is assessed based on red-yellow-black color evaluation model based on the time record of patient the healing status is analyzed.

I. INTRODUCTION

There are several problems with current practices for treating diabetic foot ulcers. First, patients must go to their wound clinic on a regular basis to have their wounds checked by their clinicians. This need for frequent clinical evaluation is not only inconvenient and time consuming for patients and clinicians, but also represents a significant health care cost because patients may require special transportation, e.g., ambulances. Second, a clinician's wound assessment process is based on visual examination. He/she describes the wound by its physical dimensions and the color of its tissues, providing important indications of the wound type and the stage of healing. Because the visual assessment does not produce objective measurements and quantifiable parameters of the healing status, tracking a wound's healing process across consecutive visits is a difficult task for both clinicians and patients.

Technology employing image analysis techniques is a potential solution to both these problems. Several attempts have been made to use image processing techniques for such tasks, including the measurement of area, or alternatively using a volume instrument system (MAVIS) or a medical digital photogram-metric system (MEDPHOS). These approaches suffer from several drawbacks including high cost, complexity, and lack of tissue classification.

To better determine the wound boundary and classify wound tissues, researchers have applied image segmentation and supervised machine learning algorithm for wound analysis. A French research group proposed a method of using a support vector machine (SVM)-based wound classification method. The same idea has also been employed in for the detection of melanoma at a curable stage. Although the SVM classifier method led to good results on typical wound images, it is not feasible to implement the training process and the feature extraction on current smartphones due to its computational demands. Furthermore, the supervised learning algorithm requires a large number of training image samples and experienced clinical input, which is difficult and costly.

Our solution provides image analysis algorithms that run on a smartphone, and thus provide a low cost and easy-to-use device for self-management of foot ulcers for patients with type 2 diabetes. Our solution engages patients as active participants in their own care, meeting the recommendation of the Committee on Quality of Health Care in America to provide more information technology solutions. The widely used commodity smartphone containing a high-resolution camera is a viable candidate for image capture and image processing provided that the processing algorithms are both accurate and well suited for the available hardware and computational resources. To convert an ordinary smartphone into a practical device for self-management of diabetic wounds, we need to address two tasks: 1) develop a simple method for patients to capture an image of their foot ulcers; and 2) design a highly efficient and accurate algorithm for real-time wound analysis that is able to operate within the computational constraints of the smartphone.

Our solution for task 1) was specifically designed to aid patients with type 2 diabetes in photographing ulcers occurring on the sole of their feet. This is particularly challenging due to mobility limitations, common for individuals with advanced diabetes. To this end, we designed and built an image capture box with an optical system containing a dual set of front surface mirrors, integrated

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

LED lighting and a comfortable, slanted surface for the patients to place their foot. The design ensures consistent illumination and a fixed optical path length between the sole of the foot and the camera, so that pictures captured at different times would be taken from the same camera angles and under the same lighting conditions. Task 2) was implemented by utilizing an accurate, yet computationally efficient algorithm, i.e., the mean-shift algorithm, for wound boundary determination, followed by color segmentation within the wound area for assessing healing status.

In our previous work, the wound boundary determination was done with a particular implementation of the level set algorithm, specifically the distance regularized level set evolution method. The principal disadvantage of the level set algorithm is that the iteration of global level set function is too computationally intensive to be implemented on smartphones, even with the narrow band confined implementation based on GPUs. In addition, the level set evolution completely depends on the initial curve which has to be predelineated either manually or by a well-designed algorithm. Finally, false edges may interfere with the evolution when the skin color is not uniform enough and when missing boundaries, as frequently occurring in medical images, results in evolution leakage (the level set evolution does not stop properly on the actual wound boundary). Hence, a better method was required to solve these problems.

To address these problems, we replaced the level set algorithms with the efficient mean-shift segmentation algorithm. While it addresses the previous problems, it also creates additional challenges, such as oversegmentation, which we solved using the *region adjacency graph* (RAG)-based region merge algorithm. In this paper, we present the entire process of recording and analyzing a wound image, using algorithms that are executable on a smartphone, and provide evidence of the efficiency and accuracy of these algorithms for analyzing diabetic foot ulcers.


This paper is organized as follows: Section II-A provides an overview of the structure of the wound image analysis software system. Section II-B briefly introduces the mean-shift algorithm used in our system and related region merge methods. Section II-C introduces the wound analysis method based on the image segmentation results including foot outline detection, wound boundary determination, color segmentation within the wound and healing status evaluation. In Section III, the GPU optimization method of the mean-shift segmentation algorithm is discussed. Section IV presents the image capture box designed for patients with diabetic foot ulcers to easily use the smartphone to take an image of the bottom of their foot. Experimental results are presented and analysed in Section V. Finally, Section VI provides an overall assessment of the wound image analysis system. A preliminary version of this paper has been reported in .

II. WOUND ANALYSIS METHOD

A. Wound Image Analysis System Overview

Our quantitative wound assessment system consists of several functional modules including wound image capture, wound image storage, wound image preprocessing, wound boundary determination, wound analysis by color segmentation and wound trend analysis based on a time sequence of wound images for a given patient. All these processing steps are carried out solely by the computational resources of the smartphone. The functional diagram of our quantitative wound assessment system is shown as in Fig. 1 and explained later. Note that the words highlighted in italics in the text correspond to specific blocks in figures with block diagrams. While the image capture is the first step in the flowchart, the image capture box is not one of the image processing steps and is therefore presented later in Section IV. A Nexus 4 smartphone was chosen due to its excellent CPU + GPU performance and high-resolution camera. Although there are likely performance variations across the cameras of modern smartphones, such a study was considered beyond the scope of this paper. After the wound image is captured, the JPEG file path of this image is added into a wound image database. This compressed image file, which cannot be processed directly with our main image processing algorithms, therefore needs to be decompressed into a 24-bit bitmap file based on the standard RGB color model. In our system, we use the built-in APIs of the Android smartphone platform to accomplish the JPEG compression and decompression task.

International Journal for Research in Applied Science & Engineering Technology (IJRASET)



Android smartphone platform to ensure high efficiency. Second, we smooth the images to remove noise (assumed mainly to be Gaussian noise produced by the image acquisition process) by using the Gaussian blur method whose standard deviation $\sigma = 0.5$ was empirically judged to be optimal based on multiple experiments.

To determine the boundary of the wound area, we first determine an outline of the foot within the image. Hence the initial *Image*

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

segmentation operation is to divide the original image into pixel groups with homogeneous color values. Specifically, the *Foot outline detection* is performed by finding the largest connected component in the segmented image under the condition that the color of this component is similar enough to a preset standard skin color. Based on the standard color checkers provided in [20], both the light and dark skin color thresholds in CIE LAB space are incorporated into the system, which means that our algorithm is expected to work for most skin colors. Afterwards, we carry out a *Wound boundary determination* based on the foot outline detection result. If the foot detection result is regarded as a binary image with the foot area marked as “white” and rest part marked as “black,” it is easy to the largest connected “black” component within the “white” part. If the wound is located at the foot region boundary, then the foot boundary is not closed, and hence the problem becomes more complicated, i.e., we might need to first form a closed boundary.

When the wound boundary has been successfully determined and the wound area calculated, we next evaluate the healing state of the wound by performing *Color segmentation*, with the goal of categorizing each pixel in the wound boundary into certain classes labeled as granulation, slough and necrosis . The classical self-organized clustering method called K-mean with high computational efficiency is used . After the color segmentation, a feature vector including the wound area size and dimensions for different types of wound tissues is formed to describe the wound quantitatively. This feature vector, along with both the original and analysed images, is saved in the result database.


The *Wound healing trend analysis* is performed on a time sequence of images belonging to a given patient to monitor the wound healing status. The current trend is obtained by comparing the wound feature vectors between the current wound record and the one that is just one standard time interval earlier (typically one or two weeks). Alternatively, a longer term healing trend is obtained by comparing the feature vectors between the current wound and the base record which is the earliest record for this patient.

B. Mean-Shift-Based Segmentation Algorithm

We chose the mean-shift algorithm, proposed in [1], over other segmentation methods, such as level set and graph cut-based algorithms, for several reasons. First, the mean-shift algorithm takes into consideration the spatial continuity inside the image by expanding the original 3-D color range space to 5-D space, including two spatial components, since direct classification on the pixels proved to be inefficient. Second, a number of acceleration algorithms are available [2]. Third, for both mean-shift filtering and region merge methods, the quality of the segmentation is easily controlled by the spatial and color range resolution parameters . Hence, the segmentation algorithm can be adjusted to accommodate different degrees of skin color smoothness by changing the resolution parameters. Finally, the mean-shift filtering algorithm is suitable for parallel implementation since the basic processing unit is the pixel.

In this case, the high computational efficiency of GPUs can be exploited.

The mean-shift algorithm belongs to the density estimation-based nonparametric clustering methods, in which the feature space can be considered as the empirical probability density function of the represented parameter. This type of algorithms adequately analyzes the image feature space (color space, spatial space or the combination of the two spaces) to cluster and can provide a reliable solution for many vision tasks . In general, the mean-shift algorithm models the feature vectors associated with each pixel (e.g., color and position in the image grid) as samples from an unknown probability density function $f(x)$ and then finds clusters in this distribution.



International Journal for Research in Applied Science & Engineering Technology (IJRASET)

Fig. 2. Mean-shift-based image segmentation sample result. (a) Original image. (b) Mean-shift-filtered image. (c) Region fused image. Note that we artificially increased the brightness and contrast of the images in this figure to highlight the over segmentation in (b) and to better observe the region fusion result in (c).

based on some rules. In the fusion step, extensive use was made of RAGs . The initial RAG was built




Fig. 5. Wound boundary determination and analysis result. (a) Foot boundary detection result. (b) Wound boundary determination result. (c) Color segmentation result within the wound boundary.





Fig. 6. Implementation flow of the mean-shift algorithm on both CPUs and GPUs.

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

After the best estimate of the wound boundary is obtained, we analyse the wound area within the boundary using a wound description model. Many methods for assessing and classifying open wounds require advanced clinical expertise and experience, and specialized criteria have been developed for diabetic foot ulcers .

The RYB (red–yellow–black) wound classification model, proposed in 1988 by Arnqvist, Hellgren and Vincent, is a consistent, simple assessment model to evaluate wounds . It classifies wound tissues within a wound as red, yellow, black or mixed tissues, which represent the different phases on the continuum of the wound healing process. Specifically, red tissues are viewed as the inflammatory (reaction) phase, proliferation (regeneration), or maturation (remodeling) phase; yellow tissues imply infection or tissue containing slough that are not ready to heal; and black tissues indicate necrotic tissue state, which is not ready to heal either . Based on the RYB wound evaluation model, our wound analysis task is to classify all the pixels within the wound boundary into the RYB color categories and cluster them. Therefore, classical clustering methods can be applied to solve this task.

For our wound image analysis, a fast clustering algorithm called K-mean is applied . K-mean is a simple unsupervised learning algorithm that solves the well-known clustering problem. A sample of the color-based wound analysis result is shown in Fig. 5(c). The results presented in Section V demonstrate the effectiveness of the K-mean algorithm for our task.

III. GPU-BASED OPTIMIZATION

Because the CPUs on smartphones are not nearly as powerful as those on PCs or laptops, an optimized parallel implementation based on GPUs is critical for the most computation-ally demanding module in the algorithm structure. For current Android-based smartphones, such as Nexus 4 from Google, the GPUs (Adreno 320) have high computational capabilities (up to 51.2 G Floating Point Operation per Second) [33]. As the experimental results in Section V shows, the hybrid implementation on both CPUs and GPUs can significantly improve the time efficiency for algorithms, which are suitable for parallel implementation.

Since our wound analysis is implemented on Android smartphones, we take advantage of the Android APIs for GPU implementations. In our case, we use the *Renderscript*, which offers a high performance computation API at the native level written in C (C99 standard) [34] and gives the smartphone apps the ability to run operations with automatic parallelization across all available processor cores. It also supports different types of processors such as the CPU, GPU or DSP. In addition, a program may access all of these features without having to write code to support different architectures or a different number of processing cores . On the Nexus 4 Android smartphone, we implement the mean-shift-based segmentation algorithm both on the Adreno 320 GPU and Quad-core Krait CPU using Renderscript. The algorithm implementation flow is shown as in Fig. 6 and is explained later. Our implementation scheme is similar to the ones used in .

The processing steps *Color space transformation*, *Color histogram generation and discretization*, and *Weight-map generation* (these steps belong to the mean-shift-filtering module introduced in Section II) are all implemented on the CPU. Afterward, all the needed data are moved to the global memory on the GPU. These data include the original image data in CIE Lab color space, the discretized color histogram for all three channels in this color space and the weight-map, which combines the edge information into the image segmentation to further improve the accuracy Because the mean-shift-based segmentation algorithm operates on each pixel of an image, and the computation, which takes place at each pixel, is independent of its distant surroundings, it is a good candidate for implementation on a parallel architecture. Hence, we developed a parallel implementation of *Mean-shift mode seeking*, which simply copies the image to the device and breaks the computation of the mode seeking into single pixels and their surrounding intermediate neighbouring region. An independent thread is spawned for the mean-shift mode seeking for each pixel. Multithreads are running at the same time on the GPU to realize the parallel computation. The number of threads running in parallel is determined by the computational capability of the GPU. In the Renderscript programming, this number is optimized automatically and does not need to be specified. After the mean-shift mode seeking, all the result modes for each pixel are moved back to the local memory of the CPU. The *Region fusion* step, as discussed in detail in is performed on the CPU.

IV. IMAGE CAPTURE BOX

Two devices for the image capture of diabetic foot ulcers have been reported in the literature. However, drawbacks to these designs are either large dimensions or high cost. Moreover, both devices require Wi-Fi connectivity and a laptop or PC for image processing.

To ensure consistent image capture conditions and also to facilitate a convenient image capture process for patients with type 2 diabetes, we designed an image capture device in the shape of a box. Hence, we term this device “the Image Capture Box.” The image capture box was designed as a compact, rugged, and inexpensive device that: 1) allows patients to both view the sole of their

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

foot on the screen of the smartphone and to capture an image since the majority of patients' wounds occur on the soles of their feet; 2) allows patients to rest their feet comfortably, without requiring angling of the foot or the smartphone camera, as patients may be overweight and have reduced mobility; and 3) accommodates image viewing and capture of left foot sole as well as right foot sole. To achieve these objectives, we make use of two front surface mirrors, placed at an angle of 90° with respect to each other, and with the common line of contact tilted 45° with respect to horizontal. A mechanical drawing of basic optical principle for foot imaging is shown in Fig. 7. The optical path is represented in blue straight lines with arrows indicating the direction.

A SolidWorks 3-D rendering of the image capture box is shown in Fig. 8. As seen in this figure, the entire box has a rectangular trapezoid shape. Rectangular openings for placing the foot and smartphone are cut into the slanted surface, shown in Fig. 8(b), which is at 45° with respect to horizontal. In this case, the patients can rest their foot comfortably and view their wounds through the smartphone camera. When using the box, the patients need to ensure that the wound is completely located




Fig. 7. Basic optical principle for foot imaging.




Fig. 8. Three-dimensional drawing of the mechanical structure of the image capture box. (a) From the back. (b) From the front. (c) Internal structure from the front.

within the opening by simply observing the image displayed on the smartphone.

To avoid the ghost image effect associated with normal back surface mirrors (reflective surface on the back side of the glass), front surface mirrors (reflective surface on the front side) are needed, as illustrated in Fig. 8(a). The optical paths for both the front

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

surface mirror and the normal mirror are shown in Fig. 9(a). As shown in Fig. 9(b) and (c), the ghost image effect has been eliminated by using front surface mirrors.

To provide consistent light for wound image viewing and capturing, warm white LED (light emitting diode) lighting, resembling daylight, is incorporated at the back side of the box. Based on our evaluation of different positions of the LED light, we found this to be the optimal location for foot imaging. The wall material for the image capture box is constructed from




Fig. 9. Ghost image caused by the normal second surface mirror. (a) Ghost image optical path; (b) ghost image using the normal mirrors for the box; (c) improvement by using the front surface mirrors.




Fig. 10. Image capture box illustration; (a) actual product of the image capture box; (b) wound image captured using the warm LED light.

¼ white acrylic where white was chosen to obtain better light reflectivity. The actual product is shown in Fig. 10(a), and a sample image captured by the image capture box is shown in Fig. 10(b).

V. EXPERIMENTAL RESULTS

A. Experimental Setups

The goal of the experimental work has been: 1) to assess the accuracy of the wound boundary determination based on the mean-shift algorithm and the color segmentation based on the K-mean algorithm; and 2) to perform an efficiency analysis by comparing the mean-shift algorithm to two other algorithms.

To test accuracy, we applied the mean-shift-based algorithm on two categories of wound images. For the first category, we used 30 images of simulated wounds, typically referred to as moulage wounds. The moulage wounds permitted us to evaluate our method under relatively consistent skin conditions and on wounds with distinct boundaries. Moulage is the art of applying mock injuries for the purpose of training emergency response teams and other medical and military personnel. In our case, we use the moulage wounds that include typical granulation, slough and necrotic tissues; the wounds were provided by *Image*

International Journal for Research in Applied Science & Engineering Technology (IJRASET)



Fig. 11. Wound images of the moulage simulation applied on the author's feet.




Fig. 12. Clinical image samples of actual patients.

Perspectives Corporation (Carson City, Nevada) and applied to the first author's foot. The selected four sample images (not all 30 images were presented in this paper) of the moulage wounds are shown in Fig. 11 and were captured with the smartphone camera placed on the image capture box.

For the second category, we evaluated our wound image analysis method on 34 images of actual patient wounds collected at the UMass-Memorial Health Center Wound Clinic (Worcester, MA), following an IRB approved protocol in accordance with Federal Regulations. The goal of selecting these typical wound images from type 2 diabetic patients is to provide a more realistic test of our wound boundary determination and color segmentation algorithms. Six selected sample images out of a total of 34 are shown in Fig. 12. Compared with the images in Fig. 11, the real wound images are more complex: they may have uneven illumination over the image plane, complex surrounding skin texture and wounds appearing in a variety of shapes, sometimes lacking distinct boundaries. Note that only two of the wound images in Fig. 12 were captured by our smartphone and image capture box; the other four images are directly from the wound image data base in UMASS Wound Clinic.

To test the algorithm efficiency, the mean-shift-based wound analysis algorithm discussed in Section II was implemented on the CPU (Quad core, 1500 MHz, Krait, 2048 MB system RAM) and the GPU (Adreno 320, Qualcomm) of the Nexus 4 Android smartphone. All the programming was done in Java in the Eclipse IDE (Integrated Development Environment).




Fig. 13. Experimental results for the images of the moulage wound simulation.

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

(a)–(d) Wound boundary determination results with mean-shift-based algorithm.

(e)–(h) Color segmentation results using the K-mean clustering method based on the mean-shift boundary determination results.

B. Experimental Results on the Images of Mouflage Wound Simulation

The wound boundary determination results of images in Fig. 11 are shown in Fig. 13. As seen from (a) to (d), the mean-shift segmentation algorithm described in Section II provides promising boundary detection results. However, there is still visibly imperfect detection as shown in Fig. 13(c) where the yellow wound tissue at the boundary has similar color to the healthy skin surrounding it.

As mentioned in Section II, after the wound boundary was determined, the K-mean color segmentation method was performed within the wound boundary. The color segmentation results are shown in (e)–(h). We assessed the color segmentation performance by comparing the original images in Fig. 11 with the results in Fig. 13. By careful observation, we found that most of the wound tissues were accurately classified. However, there were still some misclassifications in (e) where some visually light red tissues are classified as yellow tissue.

C. Experimental Results on Clinical Images From Actual Patients

The original images from actual patients, shown in Fig. 12, are more complex than is the case for the images of mouflage simulations and are taken from six different patients. The images in (a) and (b) are appropriately illuminated with a uniform skin color and well-defined wound boundaries, but the remaining four images are either not illuminated uniformly or with an uneven skin texture. The wounds in images (a), (c), and (d) are located completely within the limbs' boundary. In contrast, the wounds in images (b), (e), and (f) are located almost at the boundary.




Fig. 14. Wound boundary determination results for clinical images of real patients with preset bandwidths.

To adapt to these different conditions, we had to adjust the parameters of the algorithm for each wound image. There are three adjustable parameters in the mean-shift-based wound boundary determination algorithm for each wound image: the spatial resolution h_s , the range resolution h_r , and the fusion resolution h_f .

First, we tried parameter settings $h_s = 7$, $h_r = 6$, and $h_f = 3$ because previous work [17] showed good segmentation results with these values. Our experiments with these default setting on the real wound images in Fig. 12 did not provide satisfactory wound boundary determination results for all six images shown in Fig. 14. Note that we only try to delineate the boundary for the largest wound in the middle for the multiple-wound situation as shown in Fig. 12(b).

As mentioned in [16], only features with large spatial support are represented in the mean-shift-filtered image when h_s increased, and only features with high color contrast survive when h_r is large. In it is also stated that a larger number of regions will remain in the region fused image by employing a smaller h_f . In Section II-C, we discussed how the wound boundary determination is strongly dependent on whether a complete foot boundary can be detected. Hence, we need a better spatial resolution as well as a better region fusion resolution value to group the small regions in the foot area to a connected component when the skin texture is complex [as shown in Fig. 12(c)]. On the other hand, if the wound is located near to the foot boundary (as shown in (b), (e), and (f)), better spatial and fusion resolution is also needed to avoid the disconnected foot boundary detection, which will cause the wound boundary de-termination failure.

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

Second, we tried different parameters in specified domain ($4 \leq h_s \leq 10$, $4 \leq h_r \leq 7$, $3 \leq h_f \leq 15$) and using 0.5 as the adjustment step to customize the parameters for each wound image in Fig. 12 as shown in Table I. The corresponding wound boundary determination results are shown in Fig. 15. Much more accurate results are obtained for (b), (c), (d), and (f) by parameter adjustment.

For a more objective evaluation, we asked three experienced wound clinicians from UMass Medical School to independently label the wound area for all 34 real wound images. Then, we applied a majority vote [6], [7] for each pixel. Finally, we used the MCC (Matthews Correlation Coefficient) [39] to measure the wound boundary determination accuracy by us-

TABLE I
 PARAMETERS ADJUSTMENT FOR WOUND BOUNDARY DETERMINATION
 OF DIFFERENT IMAGES

	Image Default	Image a	Image b	Image c	Image d	Image e	Image f
Spatial resolutio n h_s	7	7	9	9	9	9	7
Range resolutio n h_r	6	6.5	6.5	8	6.5	6.5	6.5
Fusion resolutio n h_f	3	7	7	10	10	10	12




Fig. 15. Wound boundary determination results for clinical images of real patients with adjusted bandwidths.

the combined labelled wound images as the “ground truth.” The MCC returns a value between -1 (total disagreement) and $+1$ (perfect prediction). With a fixed optimal parameter setting for 34 wound images, the MCC score was 0.403. In contrast, with customized parameter settings, the MCC score improved to 0.736.

Color segmentation results provided by K-mean algorithm based on wound boundary determination results from Fig. 15 are shown in Fig. 16. The results are promising despite a small number of misclassified pixels in (c) and (f). In (c), some dark red part is recognized as black tissue. In (f), some “dark yellow” part is classified as the red tissues. In conclusion, the wound analysis task is much more complicated for the clinical images of real patients due to the complicated skin color and texture of patients’ feet, various wound locations and uneven illumination over the image plane.

D. Computational Efficiency Analysis

In this section, we present data for an efficiency analysis of: 1) the mean-shift method as compared to the level set [15] and graph cut-based algorithms [40], which are typical image segmentation algorithms; and 2) a comparison of two different implementations of the mean-shift-based method.

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

First, the three algorithms were implemented without GPU optimization on a quad-core PC with 4 GB of RAM and applied to all 64 wound images (30 moulage images and 34 real wound images). For similar segmentation results, the average computing time of level set and graph cut-based algorithms was




Fig. 16. Wound assessment results for clinical images of real patients.

approximately 15 and 10 s, respectively. In contrast, the mean-shift-based method required only about 4 s on average. Thus, the mean-shift-based wound boundary determination method provides a much better efficiency, being nearly four times faster than the level set based algorithm and twice as fast as the graph cut-based algorithm. Hence, it is reasonable to select the mean-shift-based algorithm as it delivers the most balanced performance of both boundary detection accuracy and time efficiency.

Second, we compared the mean-shift-based algorithm on the smartphone's CPU to the smartphone CPU + GPU. The average processing time of the 64 images for the CPU and CPU + GPU implementations on the smartphone is approximately 30 and 15 s, respectively. We can see that the time efficiency is significantly enhanced by GPU implementation.

While GPU + CPU implementation of the mean-shift algorithm on a laptop only provides minimal improvements in computation time over a CPU implementation, the GPU + CPU implementation on the smartphone does improve the time efficiency by about a factor of 2. This is partly because the CPU on the smartphone is not as powerful as the one on the laptop. Moreover, the Renderscript implementation utilizes both the smartphone GPU as well as the CPU and even available DSP devices on chip to provide the effective optimization.

VI. CONCLUSION

We have designed and implemented a novel wound image analysis system for patients with type 2 diabetes suffering from foot ulcers. The wound images are captured by the smartphone camera placed on an image capture box. The wound analysis algorithm is implemented on a Nexus 4 Android smartphone, utilizing both the CPU and GPU.

We have applied our mean-shift-based wound boundary de-termination algorithm to 30 images of moulage wound simulation and additional 34 images of real patients. Analysis of these experimental results shows that this method efficiently provides accurate wound boundary detection results on all wound images with an appropriate parameter setting. Considering that the application is intended for the home environment, we can for each individual patient manually find an optimal parameter setting based on a single sample image taken from the patient before the practical application. Experimental results show that a fixed parameter setting works consistently well for a given patient (the same foot and skin condition). In the future, we may consider applying machine learning approaches to enable self-adaptive parameter setting based on different image conditions. The algorithm running time analysis reveals that the fast implementation of the wound image analysis only takes 15 s on average on the smartphone for images with pixel dimensions of 816×612 .

Accuracy is enhanced by the image capture box, which is de-signed so that consistent image capture conditions are achieved in terms of the illumination and distance from camera to object. While different smartphone cameras do have slightly different color space characteristics, we have not included color calibration mainly because the most important aspect of the wound assessment system is the tracking of *changes* to the wound, both in size and color, over consecutive images captures.

Given the high resolution, in terms of pixel size, of all modern smartphone cameras, the performance of our wound analysis system is not expected to be affected by resolution differences across smartphone cameras. In fact, the original large resolution images are downsampled to a fixed spatial resolution of 816×612 pixels.

While different image noise levels for different smartphone cameras are a potential concern, we have determined, based on the experimental results, that any noise level encountered during the image capture process can be effectively removed by applying a

International Journal for Research in Applied Science & Engineering Technology (IJRASET)

Gaussian blurring filter before wound analysis.

The primary application of our wound analysis system is home-based self-management by patients or their caregivers, with the expectation that regular use of the system will reduce both the frequency and the number of wound clinic visits. One concern is that some elderly patients may not be comfortable with operating a smartphone, but this concern could be addressed by further simplifying the image capture process to a simple voice command.

An alternative deployment strategy is placing the system in wound clinics, where a nurse can perform the wound image capture and data analysis. With this implementation, the wound analysis can be moved from the smartphone to a server, which will allow more complex and computationally demanding wound boundary detection algorithms to be used. While this will allow easier and more objective wound tracking and may lead to better wound care, this implementation of the wound analysis system is not likely to reduce the number of visits to the wound clinic.

In either implementation, telehealth is an obvious extension to the wound analysis system whereby clinicians can remotely access the wound image and the analysis results. Hence, a database will be constructed on a possibly cloud-based server to store the wound data for patients.

The possibility of microbial contamination of the image capture box by the users or the environment has so far only been addressed by wiping the surface of the box with an antimicrobial wipe after each use. A better solution may be a disposable contamination barrier, which will cover the slanted surface of the box except the openings. This will avoid the patient's foot directly touching the surface of the image capture box.

The entire system is currently being used for wound tracking in the UMass-Memorial Health Center Wound Clinic in Worcester, MA. This testing at the Wound Clinic is a first step toward usability testing of the system by patients in their homes.

In future work, we plan to apply machine learning methods to train the wound analysis system based on clinical input and hopefully thereby achieve better boundary determination results with less restrictive assumptions. Furthermore, we plan to compute a healing score to be assigned to each wound image to support trend analysis of a wound's healing status.

VII. ACKNOWLEDGMENT

The authors would like to thank all the reviewers for their constructive comments which greatly improve the scientific quality of the manuscript.

REFERENCES

- [1] K. M. Buckley, L. K. Adelson, and J. G. Agazio, "Reducing the risks of wound consultation: Adding digital images to verbal reports," *Wound Ostomy Continence Nurs.*, vol. 36, no. 2, pp. 163–170, Mar. 2009.
- [2] V. Falanga, "The chronic wound: Impaired healing and solutions in the context of wound bed preparation," *Blood Cells Mol. Dis.*, vol. 32, no. 1, pp. 88–94, Jan. 2004.
- [3] C. T. Hess and R. S. Kirsner, "Orchestrating wound healing: Assessing and preparing the wound bed," *J. Adv. Skin Wound Care*, vol. 16, no. 5, pp. 246–257, Sep. 2006.