



IJRASET

International Journal For Research in
Applied Science and Engineering Technology



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 13 **Issue:** III **Month of publication:** March 2025

DOI: <https://doi.org/10.22214/ijraset.2025.67278>

www.ijraset.com

Call:  08813907089

E-mail ID: ijraset@gmail.com

A Review on Liquid Chromatography-Mass Spectrometry (LC-MS)

Surbhi Rathore

Pharmaceutical Quality Assurance

Abstract: *The hyphenation of liquid chromatography with mass spectrometry allows for the simultaneous analysis of compounds based on their retention times and mass-to-charge ratios, providing valuable information about the identity and quantity of analytes in a sample. One of the key advantages of LC-MS is its versatility, as it can be applied to a wide range of samples including biological fluids, environmental samples, pharmaceuticals, and food products. This makes it an essential technique in fields such as pharmaceutical analysis, environmental monitoring, metabolomics, proteomics, and forensic science. These techniques have become essential tools for researchers and scientists in various industries, including pharmaceuticals, environmental monitoring, and food safety. One of the most common hyphenated techniques is gas chromatography-mass spectrometry (GC-MS), which allows for the separation and identification of complex mixtures of compounds with high sensitivity and specificity. Another popular technique is liquid chromatography-mass spectrometry (LC-MS), which is widely used in drug discovery and metabolomics studies. These hyphenated techniques offer numerous advantages, such as increased sensitivity, improved selectivity, and faster analysis times. They have greatly enhanced our ability to detect trace levels of contaminants, identify unknown compounds, and quantify analytes accurately. In conclusion, the acknowledgment to LC-MS is crucial for advancing scientific research and improving our understanding of complex chemical systems. Its versatility and sensitivity make it an indispensable tool for modern analytical chemistry. Overall, introductions to LC-MS provide students with a solid foundation in this powerful analytical technique, allowing them to confidently apply it to their own research projects and experiments.*

Keywords: *Gas chromatography-mass spectrometry (GC-MS); Liquid chromatography-mass spectrometry (LC-MS); Hyphenated Technique; metabolomics; proteomics; forensic science.*

I. INTRODUCTION

Hyphenated techniques in analytical chemistry have revolutionized the field by combining two or more separation and detection methods to provide more comprehensive and accurate results. This comprehensive review will delve into the fundamentals of hyphenated techniques, their applications, advantages, and limitations (1). One of the key benefits of hyphenated techniques is their ability to separate complex mixtures with high efficiency and sensitivity. By coupling techniques such as gas chromatography-mass spectrometry (GC-MS) or liquid chromatography-mass spectrometry (LC-MS), researchers can identify and quantify a wide range of compounds in a sample (2). However, these techniques also come with challenges such as instrument complexity, cost, and data interpretation. Understanding these limitations is crucial for maximizing the potential of hyphenated techniques in analytical chemistry (3).

Hyphenated techniques in analytical chemistry have revolutionized the field by combining two or more analytical methods to provide more comprehensive and accurate results. These techniques have become essential tools for researchers and scientists in various industries, including pharmaceuticals, environmental monitoring, and food safety. One of the most common hyphenated techniques is gas chromatography-mass spectrometry (GC-MS), which allows for the separation and identification of complex mixtures of compounds with high sensitivity and specificity. Another popular technique is liquid chromatography-mass spectrometry (LC-MS), which is widely used in drug discovery and metabolomics studies. These hyphenated techniques offer numerous advantages, such as increased sensitivity, improved selectivity, and faster analysis times. They have greatly enhanced our ability to detect trace levels of contaminants, identify unknown compounds, and quantify analytes accurately (4-6). In conclusion, hyphenated techniques in analytical chemistry play a crucial role in advancing scientific research and solving complex analytical challenges. Their versatility and effectiveness make them indispensable tools for modern analytical chemists. This review provides a comprehensive overview of hyphenated techniques, highlighting their importance in modern analytical chemistry and emphasizing the need for further research and development in this area.

II. INTRODUCTIONS TO LIQUID CHROMATOGRAPHY-MASS SPECTROMETRY (LCMS)

Liquid chromatography-mass spectrometry (LC-MS) is a powerful analytical technique that combines the separation capabilities of liquid chromatography with the detection and identification capabilities of mass spectrometry. This technique has revolutionized the field of analytical chemistry by allowing scientists to separate, identify, and quantify complex mixtures of compounds with high sensitivity and specificity. The introduction to LC-MS typically involves a brief overview of the principles behind both liquid chromatography and mass spectrometry, as well as an explanation of how these two techniques are combined in LC-MS. The importance of sample preparation, instrument calibration, and data analysis are also discussed in introductory courses on LC-MS (7). Liquid chromatography-mass spectrometry (LC-MS) has revolutionized the field of analytical chemistry by providing a powerful tool for the separation, identification, and quantification of complex mixtures of compounds. This technique combines the high resolution separation capabilities of liquid chromatography with the sensitive and selective detection capabilities of mass spectrometry. The acknowledgment to LC-MS is essential in various fields such as pharmaceuticals, environmental monitoring, food safety, and forensic analysis. LC-MS has enabled researchers to detect trace levels of contaminants in water, identify metabolites in biological samples, and quantify drugs in plasma with unparalleled accuracy and precision (8-10). In conclusion, the acknowledgment to LC-MS is crucial for advancing scientific research and improving our understanding of complex chemical systems. Its versatility and sensitivity make it an indispensable tool for modern analytical chemistry. Overall, introductions to LC-MS provide students with a solid foundation in this powerful analytical technique, allowing them to confidently apply it to their own research projects and experiments.

III. PRINCIPLES OF LCMS

The principles of LC-MS involve the separation of complex mixtures into individual components using a liquid mobile phase that passes through a stationary phase. The separated compounds are then ionized and analyzed by mass spectrometry to determine their molecular weight and structure (15). LC-MS offers high sensitivity, specificity, and accuracy in identifying compounds present in a sample. It can detect trace levels of analytes even in complex matrices. Additionally, LC-MS can provide quantitative information about the concentration of compounds present in a sample. The integration of liquid chromatography (LC) and mass spectrometry (MS) techniques has revolutionized the field of analytical chemistry (16). LC-MS combines the separation capabilities of LC with the detection and identification capabilities of MS, resulting in a powerful analytical tool that is widely used in various scientific disciplines. LC separates complex mixtures into individual components based on their chemical properties, while MS identifies and quantifies these components by measuring their mass-to-charge ratios. By combining these two techniques, researchers can achieve high sensitivity, selectivity, and accuracy in their analyses (17). The integration of LC-MS has enabled scientists to analyze a wide range of compounds, from small molecules to large biomolecules such as proteins and peptides. This technique has been instrumental in drug discovery, environmental monitoring, food safety testing, and many other areas of research (18).

IV. COMBINATIONS OF HPLC AND MS

HPLC not only separates things but also provides little extra information about how a chemical might be. In fact, it is hard in HPLC to be certain about purity of a particular peak, and if it contains only a single chemical. Adding a Mass Spectrometry to this will tell you the masses of all the chemicals present in the peak, which can be used for identifying them, and an excellent method to check for the purity. Even a simple mass spec can be used as a mass-specific detector, specific for the chemical under study. More sophisticated mass detectors such as triple quadrupole and ion-trap instruments can also be used to carry out more detailed structure-dependent analysis on what is eluting off from the HPLC system (Figure 1) (35-38).




Figure 1: Workflow of Combinations of HPLC and MS

Coupling of MS to chromatographic techniques has always been desirable due to the sensitive and highly specific nature of MS compared to other chromatographic detectors. Typical LC/MS analysis begins with the liquid chromatograph (LC) separating a mixture into its chemical components. The LC pump produces a solvent stream (the mobile phase) that passes through an HPLC column (containing the stationary phase) under high pressure (39). An autosampler introduces an aliquot of sample into this solvent stream. As the solvent stream passes through the LC column, the sample separates into its chemical components. The rate at which the components of the sample elute from the column depends on their relative affinities to the mobile phase and the stationary phase. As the separated chemical components exit the LC column, they pass through a sample transfer line and enter the mass spectrometer for ionization and analysis. As the MS analyzes the ionized components and determines each mass-to-charge ratio (m/z) and relative intensity, it sends a data stream to the data system computer. In addition to supplying information about the m/z values of ionized compounds, the MS can also supply structural and quantitative information by performing MS_n experiments (40).

V. VARIOUS APPLICATIONS OF LCMS

Liquid chromatography-mass spectrometry (LCMS) is a powerful analytical technique that combines the separation capabilities of liquid chromatography with the detection and identification abilities of mass spectrometry. LCMS has a wide range of applications in various fields such as pharmaceuticals, environmental analysis, food safety, forensics, and metabolomics. In the pharmaceutical industry, LCMS is used for drug discovery and development, as well as for quality control of finished products. Environmental scientists use LCMS to detect and quantify pollutants in air, water, and soil. Food safety experts rely on LCMS to identify contaminants and ensure the safety of food products. Forensic analysts use LCMS to identify drugs, toxins, and other substances in biological samples. Overall, LCMS is a versatile tool that plays a crucial role in modern analytical chemistry and has revolutionized the way scientists analyze complex mixtures in various fields (41).

VI. CONCLUSIONS

Liquid chromatography-mass spectrometry (LC-MS) has become an indispensable tool in analytical chemistry due to its ability to separate and identify complex mixtures of compounds with high sensitivity and specificity. The hyphenation of liquid chromatography with mass spectrometry allows for the simultaneous analysis of compounds based on their retention times and mass-to-charge ratios, providing valuable information about the identity and quantity of analytes in a sample. One of the key advantages of LC-MS is its versatility, as it can be applied to a wide range of samples including biological fluids, environmental samples, pharmaceuticals, and food products. This makes it an essential technique in fields such as pharmaceutical analysis, environmental monitoring, metabolomics, proteomics, and forensic science. The future prospects for the LC-MS as a hyphenated technique are promising and exciting. As technology continues to advance, so too does the potential for this powerful analytical tool. The combination of liquid chromatography (LC) and mass spectrometry (MS) allows for the separation and identification of complex mixtures with high sensitivity and specificity. One of the key advantages of LC-MS is its ability to analyze a wide range of compounds, from small molecules to large biomolecules, making it a versatile tool in various fields such as pharmaceuticals, environmental science, and metabolomics. Additionally, advancements in instrumentation and software have improved the speed and accuracy of data analysis, further enhancing the capabilities of this technique. As researchers continue to push the boundaries of LC-MS technology, we can expect to see even greater advancements in sensitivity, resolution, and throughput. The future looks bright for LC-MS as it continues to be at the forefront of cutting-edge analytical techniques. In conclusion, LC-MS is a powerful hyphenated technique that plays a crucial role in modern analytical chemistry by providing accurate and reliable data for the identification and quantification of compounds in complex samples. Its importance cannot be overstated in advancing scientific research and improving our understanding of chemical processes.

REFERENCES

- [1] Ardrey RE. Liquid chromatography-mass spectrometry: an introduction. John Wiley & Sons; 2003 Apr 2.
- [2] Allwood JW, Goodacre R. An introduction to liquid chromatography-mass spectrometry instrumentation applied in plant metabolomic analyses. *Phytochemical Analysis: An International Journal of Plant Chemical and Biochemical Techniques*. 2010 Jan;21(1):33-47.
- [3] Niessen WM. Advances in instrumentation in liquid chromatography-mass spectrometry and related liquid-introduction techniques. *Journal of Chromatography A*. 1998 Jan 23;794(1-2):407-35.
- [4] Niessen WM. Advances in instrumentation in liquid chromatography-mass spectrometry and related liquid-introduction techniques. *Journal of Chromatography A*. 1998 Jan 23;794(1-2):407-35.
- [5] Holčápek M, Jirásko R, Lisa M. Recent developments in liquid chromatography-mass spectrometry and related techniques. *Journal of Chromatography A*. 2012 Oct 12;1259:3-15.

- [6] Trufelli H, Palma P, Famigli G, Cappiello A. An overview of matrix effects in liquid chromatography–mass spectrometry. *Mass spectrometry reviews*. 2011 May;30(3):491-509.
- [7] Keevil BG. Novel liquid chromatography tandem mass spectrometry (LC-MS/MS) methods for measuring steroids. *Best practice & research Clinical endocrinology & metabolism*. 2013 Oct 1;27(5):663-74.
- [8] Yergey AL, Edmonds CG, Lewis IA, Vestal ML. *Liquid chromatography/mass spectrometry: techniques and applications*. Springer Science & Business Media; 2013 Dec 14.
- [9] Vogeser M, Parhofer KG. Liquid chromatography tandem-mass spectrometry (LC-MS/MS)-technique and applications in endocrinology. *Experimental and clinical endocrinology & diabetes*. 2007 Oct;115(09):559-70.
- [10] Wu AH, Gerona R, Armenian P, French D, Petrie M, Lynch KL. Role of liquid chromatography–high-resolution mass spectrometry (LC-HR/MS) in clinical toxicology. *Clinical Toxicology*. 2012 Sep 1;50(8):733-42.
- [11] Wu AH, Gerona R, Armenian P, French D, Petrie M, Lynch KL. Role of liquid chromatography–high-resolution mass spectrometry (LC-HR/MS) in clinical toxicology. *Clinical Toxicology*. 2012 Sep 1;50(8):733-42.
- [12] Díaz-Cruz MS, López de Alda MJ, López R, Barceló D. Determination of estrogens and progestogens by mass spectrometric techniques (GC/MS, LC/MS and LC/MS/MS). *Journal of Mass Spectrometry*. 2003 Sep;38(9):917-23.
- [13] Covey TR, Lee ED, Bruins AP, Henion JD. Liquid chromatography/mass spectrometry. *Analytical chemistry*. 1986 Dec 1;58(14):1451A-61A.
- [14] Wu AH, French D. Implementation of liquid chromatography/mass spectrometry into the clinical laboratory. *Clinica Chimica Acta*. 2013 May 1;420:4-10.
- [15] Famigli G, Palma P, Termopoli V, Cappiello A. The history of electron ionization in LC-MS, from the early days to modern technologies: A review. *Analytica Chimica Acta*. 2021 Jul 4;1167:338350.
- [16] Tsikas D, Zoerner AA. Analysis of eicosanoids by LC-MS/MS and GC-MS/MS: a historical retrospect and a discussion. *Journal of Chromatography B*. 2014 Aug 1;964:79-88.
- [17] Barceló D, Petrovic M. Challenges and achievements of LC-MS in environmental analysis: 25 years on. *TrAC Trends in Analytical Chemistry*. 2007 Jan 1;26(1):2-11.
- [18] Niessen WM. Structure elucidation by LC-MS. Foreword. *Analisis*. 2000 Dec 1;28(10):885-7.
- [19] Korfmacher WA. Foundation review: Principles and applications of LC-MS in new drug discovery. *Drug discovery today*. 2005 Oct 15;10(20):1357-67.
- [20] Kumar PR, Dinesh SR, Rini R. LCMS—a review and a recent update. *J. Pharm. Pharm. Sci*. 2016 Mar 1;5(5):377-91.
- [21] Pitt JJ. Principles and applications of liquid chromatography-mass spectrometry in clinical biochemistry. *The Clinical Biochemist Reviews*. 2009 Feb;30(1):19.
- [22] Majors R. Current trends in HPLC column technology. *LCGC North America*. 2012 Jan 1;30(1):20-34.
- [23] Cabrera K, Lubda D, Eggenweiler HM, Minakuchi H, Nakanishi K. A new monolithic-type HPLC column for fast separations. *Journal of High Resolution Chromatography*. 2000 Jan 1;23(1):93-9.
- [24] Cruz E, Euerby MR, Johnson CM, Hackett CA. Chromatographic classification of commercially available reverse-phase HPLC columns. *Chromatographia*. 1997 Feb;44:151-61.
- [25] Majors RE. Advances in HPLC column packing design. *LC GC EUROPE*. 2003;16(6A):8-13.
- [26] Majors RE. Are you getting the most out of your HPLC column?. *LC-GC North America*. 2003 Dec 1;21(12):1124-30.
- [27] Majors R. Current trends in HPLC column usage. *LCGC North America*. 2007 Jun 1;25(6):532-44.
- [28] Jandera P, Halama M, Novotná K, Bunčková S. Characterization and comparison of HPLC columns for gradient elution. *Chromatographia*. 2003 Jan;57:S153-61.
- [29] Ali AH. High-performance liquid chromatography (HPLC): a review. *Ann. Adv. Chem*. 2022;6:010-20.
- [30] Galea C, Mangelings D, Vander Heyden Y. Characterization and classification of stationary phases in HPLC and SFC—a review. *Analytica chimica acta*. 2015 Jul 30;886:1-5.
- [31] Gowda GN, Djukovic D. Overview of mass spectrometry-based metabolomics: opportunities and challenges. *Mass Spectrometry in Metabolomics: Methods and Protocols*. 2014:3-12.
- [32] Urban PL. Quantitative mass spectrometry: an overview. *Philosophical Transactions of the Royal Society A: Mathematical, Physical and Engineering Sciences*. 2016 Oct 28;374(2079):20150382.
- [33] Trufelli H, Palma P, Famigli G, Cappiello A. An overview of matrix effects in liquid chromatography–mass spectrometry. *Mass spectrometry reviews*. 2011 May;30(3):491-509.
- [34] Feng X, Liu X, Luo Q, Liu BF. Mass spectrometry in systems biology: an overview. *Mass spectrometry reviews*. 2008 Nov;27(6):635-60.
- [35] Kebarle P. A brief overview of the present status of the mechanisms involved in electrospray mass spectrometry. *Journal of mass spectrometry*. 2000 Jul;35(7):804-17.
- [36] Vestal ML. Methods of ion generation. *Chemical reviews*. 2001 Feb 14;101(2):361-76.
- [37] McLuckey SA, Mentinova M. Ion/neutral, ion/electron, ion/photon, and ion/ion interactions in tandem mass spectrometry: do we need them all? Are they enough?. *Journal of the American Society for Mass Spectrometry*. 2011 Jan 1;22(1):3-12.
- [38] McLuckey SA, Mentinova M. Ion/neutral, ion/electron, ion/photon, and ion/ion interactions in tandem mass spectrometry: do we need them all? Are they enough?. *Journal of the American Society for Mass Spectrometry*. 2011 Jan 1;22(1):3-12.
- [39] Mei S, Chen X. Combination of HPLC–orbitrap-MS/MS and network pharmacology to identify the anti-inflammatory phytochemicals in the coffee leaf extracts. *Food Frontiers*. 2023 Sep;4(3):1395-412.
- [40] Stavrianidi A, Stekolshchikova E, Porotova A, Rodin I, Shpigun O. Combination of HPLC–MS and QAMS as a new analytical approach for determination of saponins in ginseng containing products. *Journal of Pharmaceutical and Biomedical Analysis*. 2017 Jan 5;132:87-92.
- [41] Wilson ID, Plumb R, Granger J, Major H, Williams R, Lenz EM. HPLC-MS-based methods for the study of metabonomics. *Journal of Chromatography B*. 2005 Mar 5;817(1):67-76.
- [42] Huang L, Xiao P, Guo L, Gao W. Molecular pharmacognosy. *Science China Life Sciences*. 2010 Jun;53:643-52.
- [43] Sasidharan S, Chen Y, Saravanan D, Sundram KM, Latha LY. Extraction, isolation and characterization of bioactive compounds from plants' extracts. *African journal of traditional, complementary and alternative medicines*. 2011;8(1).

- [44] Sasidharan S, Chen Y, Saravanan D, Sundram KM, Latha LY. Extraction, isolation and characterization of bioactive compounds from plants' extracts. African journal of traditional, complementary and alternative medicines. 2011;8(1).
- [45] Tetali SD, Acharya S, Ankari AB, Nanakram V, Raghavendra AS. Metabolomics of *Withania somnifera* (L.) Dunal: Advances and applications. Journal of Ethnopharmacology. 2021 Mar 1;267:113469.
- [46] Brandhorst G, Oellerich M, Maine G, Taylor P, Veen G, Wallemacq P. Liquid chromatography–tandem mass spectrometry or automated immunoassays: what are the future trends in therapeutic drug monitoring?. Clinical chemistry. 2012 May 1;58(5):821-5.
- [47] Clarke W. Immunoassays for therapeutic drug monitoring and clinical toxicology. Drug monitoring and clinical chemistry. 2004 Jan 1;5:95-112.
- [48] Baert M, Martens S, Desmet G, De Villiers A, Du Prez F, Lynen F. Enhancing the possibilities of comprehensive two-dimensional liquid chromatography through hyphenation of purely aqueous temperature-responsive and reversed-phase liquid chromatography. Analytical chemistry. 2018 Mar 18;90(8):4961-7.
- [49] Cacciola F, Mangraviti D, Mondello L, Dugo P. Hyphenations of 2D capillary-based LC with mass spectrometry. InHyphenations of Capillary Chromatography with Mass Spectrometry 2020 Jan 1 (pp. 369-412). Elsevier.
- [50] Pirok BW, Stoll DR, Schoenmakers PJ. Recent developments in two-dimensional liquid chromatography: fundamental improvements for practical applications. Analytical Chemistry. 2018 Nov 1;91(1):240-63.
- [51] Poletti A, editor. Applications of LC-MS in Toxicology. Pharmaceutical Press; 2006.
- [52] Deventer K, Pozo OJ, Verstraete AG, Van Eenoo P. Dilute-and-shoot-liquid chromatography-mass spectrometry for urine analysis in doping control and analytical toxicology. TrAC Trends in Analytical Chemistry. 2014 Mar 1;55:1-3.
- [53] Zhang A, Sun H, Wang P, Han Y, Wang X. Modern analytical techniques in metabolomics analysis. Analyst. 2012;137(2):293-300.
- [54] Zhou B, Xiao JF, Tuli L, Resson HW. LC-MS-based metabolomics. Molecular BioSystems. 2012;8(2):470-81.
- [55] Becker S, Kortz L, Helmschrodt C, Thiery J, Ceglarek U. LC–MS-based metabolomics in the clinical laboratory. Journal of Chromatography B. 2012 Feb 1;883:68-75.
- [56] Gallien S, Duriez E, Demeure K, Domon B. Selectivity of LC-MS/MS analysis: implication for proteomics experiments. Journal of proteomics. 2013 Apr 9;81:148-58.
- [57] Ishihama Y. Proteomic LC–MS systems using nanoscale liquid chromatography with tandem mass spectrometry. Journal of Chromatography A. 2005 Mar 4;1067(1-2):73-83.
- [58] Lange E, Tautenhahn R, Neumann S, Gröpl C. Critical assessment of alignment procedures for LC-MS proteomics and metabolomics measurements. BMC bioinformatics. 2008 Dec;9:1-9.
- [59] Harrieder EM, Kretschmer F, Böcker S, Witting M. Current state-of-the-art of separation methods used in LC-MS based metabolomics and lipidomics. Journal of Chromatography B. 2022 Jan 1;1188:123069.
- [60] Nygren H, Seppänen-Laakso T, Castillo S, Hyötyläinen T, Orešič M. Liquid chromatography-mass spectrometry (LC-MS)-based lipidomics for studies of body fluids and tissues. Metabolic Profiling: Methods and Protocols. 2011:247-57.



10.22214/IJRASET



45.98



IMPACT FACTOR:
7.129



IMPACT FACTOR:
7.429



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089  (24*7 Support on Whatsapp)