



IJRASET

International Journal For Research in
Applied Science and Engineering Technology



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 8 Issue: V Month of publication: May 2020

DOI: <http://doi.org/10.22214/ijraset.2020.5020>

www.ijraset.com

Call:  08813907089

E-mail ID: ijraset@gmail.com

Cough Monitoring with Cardiopulmonary Measurements

Ms. E. Niranjana¹, S. Glory Sahana², M. Devitha³, M. Sivasankari⁴

¹Assistant Professor/Dept of BME, ^{2,3,4}BioMedical Engineering, Rajiv Gandhi College of Engineering & Technology, Puducherry, India

Abstract: An independent cough checking device is a gadget that permits quantitative cough estimations in the patients. The gadget is being practiced by two objectives: comfort and security. Internet of Things(IoT) which is a network based on physical system in which it can be exhibited in the form of typical embedded system including electronic devices. The assessment of its power and recurrence that furnishes clinical data of patient with chronic ceaseless cough. The recognition calculation pursues a methodology with accounts of cough sounds. The normal recognition rate was 82% at a normal caution pace of occasions when considering the vitality edge to every normal patient. This device utilizes sound chronicles from a smaller than normal receiver to distinguish the cough sound and it utilizes respiratory sensors to detect the in and out progression of respiratory system. Heart sensor is utilized to identify the pulse estimations during the procedure. The impact of nonstop cough might be hurtful to the patient by meddling with breathing, rest, muscle throb, sleep insomnia and fatigue. Over 90% of COPD patients happen in low and center pay nations. By breaking down the signal, we can recognize typical and irresistible cough and fast restorative consideration is given to the individuals experiencing respiratory issue. The outcome recommend that can be applied to the detection of cough sounds and could permit the dismissal of a portion of the inaccurately identified devices. With the help of these sensors the data has been collected and then it is uploaded in the cloud for further analysis.

Keywords: Cough, COPD, Heart, Respiration, Sensor, cloud

I. INTRODUCTION

Cough is a normal protective reflex which clears the respiratory tract and prevents the entrance of noxious material in to the respiratory system. Chronic obstructive pulmonary disease, the patients with COPD frequently complain of cough that in turns, is associated with an increased risk of frequent hospitalizations. Cough analysis is normally performed by means of manual cough checking yet the requirement for long haul and the expanding number of COPD patients have inspired towards new approaches for programmed cough examination. The cough which include an initial inspiration, glottal closure and development of high thoracic pressure, followed by an expiration when the glottis open. The cough receptors are located in the central airways in brain. Sound is generated during the cough due to airflow vibration of tissues and movement of fluid in the lungs. The most common cause of disease is cough. Sometimes allergies can cause cough which causes a dry or dicky cough. These diagnostic approaches on chronic cough has been done so far to the limit occurrence of cough related disease. The pulmonary diseases have varying cough pattern in which all seem to share cough as common symptom. A reliable measure of cough is needed so that the severity of cough in a particular patient and the effectiveness of treatment can be assessed. The basic requirements for an automatic cough monitor are the possibility to record over a representative amount of time using a portable recording system and it detects the occurrence of cough sound from the recordings using a patterned algorithm. A reliable cough detection algorithm to could reduce the amount of data to be manually analyzed and it improves the usefulness of cough monitors. The variability of cough sounds between different subjects and occurrence of similar sounds in the recordings that can be taken by the algorithm as being cough sounds (speech, laughing, sneezing) and other sounds. Hence a low-cost cough monitoring system can be designed for classification of various lung disease using these sensors (heart, respiratory, cough) and analysing the signal pattern for earlier screening of pulmonary diseases.

II. OBJECTIVE

The main aim of our project is for the people who cannot afford the expensive and daily check-up for their health. For this purpose, introduced self-contained smart health ambulatory monitoring system which is non-invasive and it give easy and assured caring unit is proposed and it is preferred to monitor cough continuously for long periods. The monitoring of cough for extended periods of time has long been recognized as an important step towards a better understanding of this symptoms, and a better management of chronic cough patients. This system reduces time with safely handled equipment. It is demonstrated significantly rater overall reduction in cough bouts, cough components, and cough effort and an increase in cough latency for patients.

III. EXISTING SYSTEM

The algorithm comprises of three principle squares to perform programmed cough identification, cough characterization and whooping sound detection. Each of the extract relevant features from the audio signal and subsequently classify them using a logistic regression model. The yield from these blocks is ordered to give the finding of pertussis probability. The performance of the proposed algorithm is evaluated using audio recordings from patients. The algorithm is able to diagnose all pertussis successfully from all audio recordings with any false diagnosis. It can also automatically detect individual cough sounds with an accuracy of 92% and PPV of 97%. The low multifaceted nature of the proposed algorithm combined with its high precision exhibits that it very well may be promptly diploid utilizing smart phones and can be incredibly valuable for snappy recognizable proof of prior screening pertussis and for disease outbreak control.




Fig. 1 existing model

An algorithm for automated diagnosis of pertussis using audio signals by analyzing cough and whoop sound is presented.

IV. PROPOSED SYSTEM

In this paper detection of cough events is introduced and it uses hidden Markov models which are trained for patients. The overall device is automated as an operator must identify cough fractions of the recorded sound. The technologies through which integration can be formed into an elegant health care patient statistic monitoring system.

The projected system which can transfer the patient data in a manner safe sound. We introduced heart rate sensor which is vital health parameter that is related to the soundness of the human cardio vascular system. This vacillation of blood can be distinguished through an optical detecting system. This computerized yield can be associated with microcontroller straightforwardly to gauge the beats per minute (BPM) rate. It works on light modulation by blood flow through each pulse.

A flow sensor is used to sense the localized element that measures displacement variations induced by inhaling or exhaling. This airflow sensor is a device used to monitor airflow rate of a patient in need of respiratory help and measures the peak of the patients breathing, or the inflection point between inhalation and exhalation.

Sound sensor can detect a sound (cough) intensity of the patient. The main component of the module is simple microphone which is based on a amplifier. In this system heart rate, breath (flow sensor) and cough (sound sensor) is measured for the patient health unit.




Fig. 2 block diagram

It is used for the smart health monitoring system. This cough acoustic waveform is used to computerized system that quantifies the energy expanded in a cough by digitizing the audio signal and measuring the time under the transform. The graph shows that the computerized cough acquisition and analyse system are a valid and methodology for evaluating cough associated patients. In this graph analysis shows the fact that patient experiencing acute cough demonstrate a large variability in their tussises profile over time, probably as a result of disease fluctuation can be obtained when the cough acoustic waveform signal differs.




Fig. 3 acoustic wavelength of cough

V. DESCRIPTION OF SCHEME

A. Sensor

A sensor a device that sees and react to a type of commitment from the physical condition. In this paper we present three sensors. The sensors are heart beat estimating sensor (KY - 039), Microphone sound sensor (VMA309), Flow sensor (POW110D3B).

- 1) **Arduino Microcontroller(ATMEGA 328):** ATMEGA is a single chip microcontroller made by Atmel in the family megaAVR. It has an altered equipment engineering. It is an 8-piece and 28 pins AVR Microcontroller, pursues RISC Architecture and it has program memory of 32KB. The Atmega 328 is one of the microcontroller chips are used with the popular Arduino Duemilanove sheets. It goes about as an interface. Outer (non-USB) power can come either from AC to DC connector (divider mode) or battery. The connector can be related by halting 2.1mm center positive fitting into the boards ability jack. Leads from the battery can be implanted in the Gnd and vin stick header of the power connector.



Fig. 3 arduino microcontroller

- 2) **Heart Beat Sensor:** The KY – 039 heart beat estimating sensor module utilizes splendid infrared (IR) LED and a phototransistor to distinguish the beat of the finger, a red LED flashes with each heartbeat. The LED is the light side of the finger, and the phototransistor on the opposite side of the finger, the phototransistor used to acquire a transition radiated when the circulatory strain beat by the finger, when the obstruction of the transistor must be marginally changed. We pick a high restriction resistor R1 that most of the light through the finger is consumed, it is appealing that the photo transistor is delicate enough. Opposition can be kept to shield stray light into the phototransistor. The light will vacillate from 50HZ to 60HZ.



Fig. 4 heart sensor

- 3) *Flow Sensor*: Flow sensor consist of a valve body, and a halleffect sensor. When air flows through the rotor the rotor rolls. Its speed changes with different rate of flow. The halleffect sensor output in corresponding pulse signal. The architecture 8-bit RISC processor core. It joins 32KB ISP streak memory with read while compose abilities. A programmable watch dog timer is fixed with internal oscillator. The most extreme working frequency of ATMEGA 328 is 20MHz.It consist of two external interrupts. ATMEGA 328 is commonly known as Pico power. It utilizes a basic turning wheel that heartbeats the lobby impact sensor. The working voltage of sensor is 5 – 24V. The working weight of sensor is under 1.2Mpa.



Fig. 5 flow sensor

- 4) *Sound Sensor*: This sound locator module recognizes sound somewhere in the range of 48 and 66db and has simple has too advanced yield. It is commonly used to distinguish the sound affectability and the info is typically intensified. It makes a constant yield voltage sign of the receiver. It is utilized to change over the acoustic to electric transducer that distinguishes the sound sign and changes over it into electrical sign. It can without much of a stretch incorporate with rationale module on the info side of the circuit. It is additionally called sound-related sensors. It is commonly utilized for identifying sound powers predominantly used to separate hack sound and other sound. It uses a receiver which supplies the commitment to a speaker, crest pointer and cradle. At the point when the sensor recognizes thee sound, it forms a yield signal voltage which is sent a microcontroller then performs fundamental preparing.



Fig. 6 sound sensor

VI.RESULTS AND DISCUSSION

The breathing movement examination including the respiratory recurrence, with the discovery of each hack occasion are exceedingly significant respiratory parameters that can help being developed of checking advances. While the framework can distinguish the hacking event, progressively complex calculation have been proposed in discourse recognition and that can be utilized to improve the effectiveness power.

At the same time this process produces the output simultaneously for cardiac rate, cough rate and pulmonary respiration rate. The measured sensor values are continuously uploaded in the cloud server through which the patient health condition can be monitored at anywhere by the caretakers.



Fig. 7 hardware output




Fig. 8 heart rate graph




Fig. 9 flow rate graph




Fig. 10 cough rate graph

VII. CONCLUSIONS

As we proposed a brilliant individual consideration framework to help the patients in respects of the change in natural conditions they exist. In this framework we gauged pulse, breathe (stream sensor) and hack (sound sensor) focus. The framework gives solace to the client with its little size circuit, a lot faster and simpler to utilize.

REFERENCES

- [1] Q. Jian, J. Yang, Y. Yu, P. Bjorkholm, and T. McKelvey, "Detection of breathing and heartbeat by using a simple uwb radar system", 8th European Conference on, April 2014, pp.3078-3081.
- [2] A. Bates, M.J. Ling, J. Mann, and D.K Arvind, "Respiratory rate and flow wave for estimation from tri-axial accelerometer data", 2010 International Conference on Body Sensor Networks, June 2010, pp.144-150.
- [3] T. Drugman, J. Urbain, N. Bauwens, R. Chessini, C. Valderrama, P. Lebecque, and T. Dutoit, "Objective study of sensor relevance for automatic cough detection", IEEE journal of Biomedical and Health Informatics, vol.17, no.3, pp. 11493-1501.
- [4] J. Smith, "Monitoring chronic cough: current and future techniques", Expert Rev Resp Med, vol.4, no.5, pp.673-683, Oct 2010.
- [5] M. G. Crooks et al., "Objective measurement of cough frequency during COPD exacerbation convalescence," Lung, vol. 194, no. 1, pp. 117-120, 2016.
- [6] P.M. Calverley, "COPD: early detection and intervention," CHEST Journal, vol. 117, no. 5_suppl_2, pp. 365S-37 IS, 2000.
- [7] W. Li, B. Tan, and R.J. Piechocki, "Non-contact breathing detection using passive radar," in 2016 IEEE International Conference on Communications (ICC), May 2016, pp. 1-6.
- [8] B. Ferdousi, S.M.F. Ahsanullah, K. Abdullah-Al-Mamun, and M.N. Huda, "Cough detection using speech analysis," in 2015 18th International Conference on Computer and Information Technology (ICCIT).



10.22214/IJRASET



45.98



IMPACT FACTOR:
7.129



IMPACT FACTOR:
7.429



INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089  (24*7 Support on Whatsapp)